



Certificate of Compliance

Certificate: NA201911131

Date Issued: October 21, 2019

Project: 369092-1.2

Issued to: Velp Scientifica S.r.l.
Via Stazione 16,
20865 Usmate (MB)
Italy

The products listed below have been certified as being compliant with all applicable requirements of the specifications listed and are eligible to bear the following certification mark



Authorized by:

S.C. Beck, Director of Certification

PRODUCTS

MEASUREMENT, CONTROL, OR LABORATORY EQUIPMENT - Certified to US and Canadian Standards

Product: Motor controller

Model: F201A0500xx; F201A0505xx (In the model name "xx" is composed by two alphanumeric symbols or blank and indicates only a different graphic front panel. Not safety related).

Ratings: 115-230 V~, 50-60 Hz, 300 W, Class I equipment (both models)

APPLICABLE REQUIREMENTS

UL Std. No. 61010-1 3rd Edition - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-1-12 Third Edition – Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements

This certificate is issued on condition that the holder complies and will continue to comply with the requirements of the above mentioned specifications and pursuant to the terms and conditions specified in the Certification Agreement.



Supplement to Certificate of Compliance

Certificate: NA201911131

Project: 369092-1.2

Customer Number: 101740

Nemko North America grants a license to the applicant to apply the Certification Mark to the certified products and that the mark shall only be affixed at the following factory locations

Factory Information

Factory Name	Factory Number	Location
Velp Scientifica S.r.l.	101740	Via Stazione 16, 20865 Usmate (MB) Italy

Product Certification History

Project	Date	Description
369092-1.1	October 18, 2019	Original Certification: Model: F201A0500xx; F201A0505xx (In the model name "xx" is composed by two alphanumeric symbols or blank and indicates only a different graphic front panel. Not safety related). Ratings: 115-230 V~, 50-60 Hz, 300 W, Class I equipment (both models)
369092-1.2	October 21, 2019	Update: Change accessories code from 40002633 to A00000003+A00000002 on rear side of appliance inlet on page 6 of description report; Change board code from 40002608 - 40002704 to 40002704 for photographs of model F201A0500xx on page 37 of description report.

This Supplement forms an integral part of the Certificate of Compliance

