



Certificate of Compliance

Certificate: NA201911126

Date Issued: July 13, 2020

Project: 397689-10.1

Issued to: Velp Scientifica S.r.l.
Via Stazione 16,
20865 Usmate (MB)
Italy

The products listed below have been certified as being compliant with all applicable requirements of the specifications listed and are eligible to bear the following certification mark



Authorized by:

Marilyn Laroche, Technical Assessor

PRODUCTS

MEASUREMENT, CONTROL, OR LABORATORY EQUIPMENT - Certified to US and Canadian Standards

Product: Overhead Stirrer

Model: Code F201w04yzxx (See description report for explanation of letters w, y, z and xx).

Ratings: 190 W 230 V~ 50-60 Hz Class I equipment (for models with Code F201004yzxx); 190 W 115 V~ 60 Hz Class I equipment (only for model with Code F201104yzxx)

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No.61010-1-12 + Amd.1	– Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements
UL Std. No. 61010-1 3rd Edition +Amd.1	–Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements

Supplemental standards included in evaluation:

CSA C22.2 No. 61010-2-051:15	Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 2-051: Particular requirements for laboratory equipment for mixing and stirring (Adopted IEC 61010-2-051:2015, Third edition, 2015-03, with Canadian deviations)
UL Std.No. 61010-2-51 3rd Edition	Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use - Part 2-051: Particular Requirements for Laboratory Equipment for Mixing and Stirring

This certificate is issued on condition that the holder complies and will continue to comply with the requirements of the above mentioned specifications and pursuant to the terms and conditions specified in the Certification Agreement.



Supplement to Certificate of Compliance

Certificate: NA201911126

Project: 397689-10.1

Customer Number: 101740

Nemko North America grants a license to the applicant to apply the Certification Mark to the certified products and that the mark shall only be affixed at the following factory locations

Factory Information

Factory Name	Factory Number	Location
Velp Scientifica S.r.l.	101740	Via Stazione 16, 20865 Usmate (MB) Italy

Product Certification History

Project	Date	Description
342099-1.1	October 03, 2019	Original Certification: Model: OHS 200 Advance; OHS 100 Advance; OHS 60 Advance; OHS 200 Digital; OHS 100 Digital; OHS 60 Digital; OHS 40 Digital; OHS 20 Digital; Code F201w04yzxx (See description report for explanation of letters w, y, z and xx). Ratings: 190 W 230 V~ 50-60 Hz Class I equipment (for models with Code F201004yzxx); 190 W 115 V~ 60 Hz Class I equipment (only for model with Code F201104yzxx)
342099-1.2	October 15, 2019	Update: Addition of missing label "OHS 100 Advance"; Correction of Display Graphic board code from 100072067 to 40002608
397689-10.1	July 13, 2020	Update: - addition of a new alternative motor (mod. FL57BLSH60-325V-4000A manufactured by Fulling motor) dedicated only for models OHS 40 Digital and OHS 20 Digital (both 230 V~ and 115 V~ models); - updated of Canada national differences of CAN/CSA-C22.2 No.61010-1 + Amd. 1 - updated of U.S.A. national differences of UL 61010-1 (3rd Ed.); Am. 1 - introducing of alternative components for M1, F1, F2, V1, C6, C7, C8, C5, C9, F3, U1, J8, J9 - other minors changes. - correct model name to Code F201w04yzxx

This Supplement forms an integral part of the Certificate of Compliance

